

中环科仪 5156 高阻基准指标

硬件指标

	基本误差	最大耐压	温漂
100M	1%	10kV	25ppm/°C
1Gohm	1%	5kV	25ppm/°C
10Gohm	0.5%	7.5kV	100ppm/°C
100Gohm	1%	7.5kV	100ppm/°C
1000pF	5%	250V	C0G 标准 30ppm/°C
100nF	5%	100V	C0G 标准 30ppm/°C

和 datron 4000 联合调试

2nA	1Gohm	+2nA	1.9956099V
	1Gohm	-2nA	-1.9959099V
20nA	100M	+20nA	1.9945099V
	100M	-20nA	-1.9945099V
200nA	100M	+200nA	19.945099V
	100M	-200nA	-19.945099V
2μA	100M	+2μA	+199.49099V
	100M	-2μA	-199.48899V
200pA	10G	+200pA	+2.006999V
	10G	-200pA	-2.006999V
20pA	100G	+20pA	1.9953599V
	100G	-20pA	-1.9933599V